UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/696,845	10/29/2003	Stephen P. Mangin	A-72194/ENB	9143
32940 7590 08/31/2007 DORSEY & WHITNEY LLP 555 CALIFORNIA STREET, SUITE 1000 SUITE 1000 SAN FRANCISCO, CA 94104			EXAMINER	
			PRONE, CHRISTOPHER D	
			ART UNIT	PAPER NUMBER
			3738	
			MAIL DATE	DELIVERY MODE
			08/31/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)			
	10/696,845	MANGIN ET AL.			
Office Action Summary	Examiner	Art Unit			
	Christopher D. Prone	3738 ··			
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply					
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DOWN THE MAILING DOWN THE MAILING DOWN THE MAILING DOWN THE MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period to Failure to reply within the set or extended period for reply will, by statute Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be timwill apply and will expire SIX (6) MONTHS from a cause the application to become ABANDONE	L. lely filed the mailing date of this communication. D (35 U.S.C. § 133).			
Status					
 Responsive to communication(s) filed on 18 July 2007. This action is FINAL. This action is non-final. Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. 					
Disposition of Claims					
 4) Claim(s) 1-17 and 23-34 is/are pending in the application. 4a) Of the above claim(s) 5-7,15-17,24,25,28,29,31,32 and 34 is/are withdrawn from consideration. 5) Claim(s) is/are allowed. 6) Claim(s) 1-4,8-14,23,26,27,30 and 33 is/are rejected. 7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/or election requirement. 					
Application Papers					
9) The specification is objected to by the Examine 10) The drawing(s) filed on is/are: a) acc Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the Example 11.	epted or b) objected to by the Education of the Education of the drawing (s) be held in abeyance. See tion is required if the drawing (s) is obj	e 37 CFR 1.85(a). ected to. See 37 CFR 1.121(d).			
Priority under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 					
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail Da 5) Notice of Informal P 6) Other:	te			

Application/Control Number: 10/696,845

Art Unit: 3738

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) The invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-4, 8, 12-14, 23, 26, 27, 30, 33, and 35 are rejected under 35 U.S.C. 102(b) as being anticipated by United States Patent 5,653,748 Strecker.

Strecker discloses the same invention being a prosthetic delivery device (10) comprising a flexible elongated member (11) having proximal and distal ends, a means for releasably securing an expandable stent prosthesis (15). The device further comprises a crocheted material (14) including a thread having a plurality of loops, a trigger pull release (24), and a visual marker comprising a colored loop (20) extending around the flexible member. The last loop 20 disclosed by Strecker is visually distinct from the remainder of the loops because of its crocheted looped shape. Each loop has a clear beginning and end located further down the prosthesis thereby making them each visually distinguishable by the operator.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

Application/Control Number: 10/696,845

Art Unit: 3738

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 9-11 are rejected under 35 U.S.C. 103 as being unpatentable over United States Patent 5,653,748 Strecker in view of United States Patent 5,480,423 Ravenscroft.

Strecker discloses the invention substantially as claimed being described above. However, Strecker does not disclose that his prosthesis expands to a larger radius and a shorter length upon implantation

Ravenscroft teaches the use of a prosthetic delivery device comprising a length shortening self-expanding stent in the same field of endeavor for the purpose of providing a concentrated expansion force within a body lumen, shown best in figures 2a – 2f.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to combine the stent of Ravenscroft with the delivery device of Strecker in order to provide a concentrated expansion force within a body lumen, which can be accurately tracked by the operator throughout the insertion process.

In regards to claim 10 the claim recites that the visual marker only needs to be attached to "one of" the distal extremity of the flexible elongate member and the prosthesis a distance from one of the first and second ends of the prosthesis equal to the second length. In view of this broad limitation Strecker discloses a visual marker attached to the distal end of the flexible member. By meeting the criteria of the first option the location of the marker with respect to the shortening of the stent is irrelevant.

Art Unit: 3738

Response to Arguments

Applicant's arguments filed 7/18/07 have been fully considered but they are not persuasive.

Applicant argues that neither Strecker nor Ravenscroft disclose a visual marker overlying the stent intermediate the first and second ends. However the visual marker of Strecker is located on top of the stent and extending between the distal and proximal most ends of the stent as shown in figure 1. In view of the amendments the visual markers of Ravenscroft are no longer used in the combination disclosed above.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3738

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Prone whose telephone number is (571) 272-6085. The examiner can normally be reached on Monday Through Fri 8:30 to 5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Christopher D Prone Examiner Art Unit 3738

CDP

CORRINE McDERMOTT
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700